

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS083118\
 Data File : PS001665.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2018 22:05
 Operator : EI\SJ
 Sample : J4725-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 CONCRETE-BEDROCK

Manual Integrations
 APPROVED

Sohil
 9/4/2018 1:12:44 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 01 00:24:25 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082218.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 23 02:55:24 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	11.486	10.986	98130901	94013854	531.585m	464.197m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS083118\
Data File : PS001665.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2018 22:05
Operator : EI\SJ
Sample : J4725-01
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
CONCRETE-BEDROCK


Manual Integrations
APPROVED

Sohil
9/4/2018 1:12:44 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 01 00:24:25 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS082218.M
Quant Title : 8080.M
QLast Update : Thu Aug 23 02:55:24 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA

R.T.: 11.486 min
Delta R.T.: -0.006 min
Response: 98130901
Conc: 531.59 ng/ml m

Instrument :
ECD_S
ClientSampleId :
CONCRETE-BEDROCK

Manual Integrations
APPROVED

Sohil
9/4/2018 1:12:44 PM



#4 2,4-DCAA

R.T.: 10.986 min
Delta R.T.: -0.003 min
Response: 94013854
Conc: 464.20 ng/ml m