

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101625\
 Data File : PS032073.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Oct 2025 21:08
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 10/17/2025
 Supervised By : Sohil Jodhani 10/17/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 17 04:02:46 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101325.M
 Quant Title : 8080.M
 QLast Update : Mon Oct 13 12:51:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.281	7.745	3616.4E6	846.0E6	758.626	768.842
Target Compounds							
1) T	Dalapon	2.660	2.688	4876.3E6	2498.6E6	631.034	664.824
2) T	3,5-DICHL...	6.448	6.697	5013.4E6	1474.5E6	712.245	706.884
3) T	4-Nitroph...	7.098	7.303	942.7E6	977.6E6	706.282m	644.253
5) T	DICAMBA	7.463	7.939	15005.7E6	6192.3E6	716.521	714.974
6) T	MCPD	7.642	8.036	910.9E6	180.8E6	74.491	72.853
7) T	MCPA	7.791	8.284	1109.2E6	278.1E6	75.375	73.722
8) T	DICHLORPROP	8.171	8.659	3331.0E6	1371.3E6	700.651	700.242
9) T	2,4-D	8.411	9.004	3525.9E6	1457.2E6	726.117	701.152
10) T	Pentachlo...	8.708	9.507	49047.5E6	39718.0E6	730.749	766.403
11) T	2,4,5-TP ...	9.280	9.892	20348.2E6	14419.3E6	720.495	718.275
12) T	2,4,5-T	9.577	10.323	18990.9E6	12488.8E6	752.287	728.436
13) T	2,4-DB	10.154	10.892	2362.8E6	887.8E6	762.121	687.082
14) T	DINOSEB	11.345	11.264	15745.6E6	10973.1E6	706.054	721.769
15) T	Picloram	11.182	12.390	16525.5E6	21690.8E6	700.892	651.533
16) T	DCPA	11.643	12.305	26284.4E6	23639.8E6	715.413	711.793

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS101625\
Data File : PS032073.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Oct 2025 21:08
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 10/17/2025
Supervised By : Sohil Jodhani 10/17/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 17 04:02:46 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS101325.M
Quant Title : 8080.M
QLast Update : Mon Oct 13 12:51:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

