

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111124\
 Data File : PS028442.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2024 06:59
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 11/12/2024
 Supervised By :Ankita Jodhani 11/12/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 06:56:47 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110924.M
 Quant Title : 8080.M
 QLast Update : Fri Nov 08 17:46:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.242	7.748	2061.2E6	1082.1E6	746.624	679.113
Target Compounds							
1) T	Dalapon	2.642	2.702	2383.9E6	1734.7E6	668.112	611.757
2) T	3,5-DICHL...	6.413	6.700	2794.7E6	1490.8E6	698.443	629.290
3) T	4-Nitroph...	7.041	7.274	1223.7E6	631.0E6	659.959	595.862m
5) T	DICAMBA	7.430	7.949	8699.5E6	4697.7E6	722.504	655.236
6) T	MCP	7.614	8.052	559.5E6	367.6E6	68.412	62.568
7) T	MCPA	7.765	8.296	798.0E6	494.9E6	71.172	61.809
8) T	DICHLORPROP	8.139	8.665	2200.0E6	1125.0E6	709.754	631.599
9) T	2,4-D	8.370	8.995	2479.6E6	1263.6E6	705.607	653.430m
10) T	Pentachlo...	8.670	9.525	34160.5E6	16629.5E6	758.413	667.401
11) T	2,4,5-TP ...	9.249	9.902	13794.5E6	6936.6E6	728.750	654.876
12) T	2,4,5-T	9.542	10.323	14081.8E6	6847.6E6	726.579	659.948
13) T	2,4-DB	10.118	10.890	2045.0E6	856.7E6	661.619	646.397m
14) T	DINOSEB	11.332	11.270	11612.4E6	4644.6E6	716.221	647.114
15) T	Picloram	11.137	12.367	22398.0E6	9477.2E6	713.176	658.485
16) T	DCPA	11.624	12.314	19474.2E6	7437.1E6	728.649m	663.424

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS111124\
Data File : PS028442.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2024 06:59
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDCCC750

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 11/12/2024

Supervised By :Ankita Jodhani 11/12/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 06:56:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS110924.M
Quant Title : 8080.M
QLast Update : Fri Nov 08 17:46:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

