

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121919\
 Data File : PS008111.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2019 19:14
 Operator : SM\AJ
 Sample : K6299-08
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 05:37:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121819.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 18 11:09:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.442	5.174	191.2E6	88676084	127.631	120.823
Target Compounds							
1) T	Dalapon	2.000	1.672	69469703	42287093	37.109	47.740 #
2) T	3,5-DICHL...	0.000	4.505	0	67869578	N.D.	71.974
5) T	DICAMBA	5.533	0.000	96280298	0	17.395	N.D. #
6) T	MCP	0.000	5.407	0	10726193	N.D.	5.489
7) T	MCPA	5.906	5.541	35909213	32889393	6.291	11.013 #
8) T	DICHLORPROP	0.000	5.838	0	14994550	N.D.	21.861
9) T	2,4-D	0.000	6.075	0	1742838	N.D.	2.142
10) T	Pentachlo...	0.000	6.385	0	29989830	N.D.	2.948
11) T	2,4,5-TP ...	0.000	6.773	0	26216198	N.D.	6.481
14) T	DINOSEB	0.000	7.960	0	37247940	N.D.	14.663
15) T	Picloram	0.000	8.939	0	15029896	N.D.	2.497



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121919\
Data File : PS008111.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2019 19:14
Operator : SM\AJ
Sample : K6299-08
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 05:37:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121819.M
Quant Title : 8080.M
QLast Update : Wed Dec 18 11:09:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.000 min
Delta R.T.: -0.006 min
Response: 69469703
Conc: 37.11 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 1.672 min
Delta R.T.: -0.046 min
Response: 42287093
Conc: 47.74 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 4.800 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 4.505 min
Delta R.T.: 0.043 min
Response: 67869578
Conc: 71.97 ng/ml



#4 2,4-DCAA


R.T.: 5.442 min
Delta R.T.: 0.020 min
Response: 191205198
Conc: 127.63 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 5.174 min
Delta R.T.: 0.021 min
Response: 88676084
Conc: 120.82 ng/ml




#5 DICAMBA

R.T.: 5.533 min
Delta R.T.: -0.039 min
Response: 96280298
Conc: 17.39 ng/ml



#5 DICAMBA


R.T.: 0.000 min
Exp R.T.: 5.277 min
Response: 0
Conc: N.D.



#6 MCPP


R.T.: 0.000 min
Exp R.T.: 5.722 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#6 MCPP

R.T.: 5.407 min
Delta R.T.: 0.046 min
Response: 10726193
Conc: 5.49 ug/ml




#7 MCPA

R.T.: 5.906 min
Delta R.T.: 0.054 min
Response: 35909213
Conc: 6.29 ug/ml



#7 MCPA


R.T.: 5.541 min
Delta R.T.: 0.005 min
Response: 32889393
Conc: 11.01 ug/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 6.193 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 5.838 min
Delta R.T.: 0.033 min
Response: 14994550
Conc: 21.86 ng/ml




#9 2,4-D

R.T.: 0.000 min
Exp R.T.: 6.409 min
Response: 0
Conc: N.D.



#9 2,4-D


R.T.: 6.075 min
Delta R.T.: 0.000 min
Response: 1742838
Conc: 2.14 ng/ml



#10 Pentachlorophenol


R.T.: 0.000 min
Exp R.T.: 6.663 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#10 Pentachlorophenol

R.T.: 6.385 min
Delta R.T.: -0.051 min
Response: 29989830
Conc: 2.95 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T.: 7.195 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 6.773 min
Delta R.T.: -0.025 min
Response: 26216198
Conc: 6.48 ng/ml



#14 DINOSEB


R.T.: 0.000 min
Exp R.T. : 9.075 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 7.960 min
Delta R.T.: 0.013 min
Response: 37247940
Conc: 14.66 ng/ml



#15 Picloram

R.T.: 0.000 min
Exp R.T. : 8.912 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 8.939 min
Delta R.T.: 0.057 min
Response: 15029896
Conc: 2.50 ng/ml