

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121919\
 Data File : PS008106.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2019 17:25
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 12/20/2019 8:08:12 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 05:35:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121819.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 18 11:09:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.441	5.170	1143.3E6	560.1E6	763.188	763.129
Target Compounds							
1) T	Dalapon	2.024	1.735	1091.9E6	523.5E6	583.241	590.987
2) T	3,5-DICHL...	4.817	4.480	1373.6E6	665.4E6	701.215	705.618
3) T	4-Nitroph...	5.302	4.864	409.6E6	248.2E6	646.123	635.250
5) T	DICAMBA	5.588	5.294	4101.2E6	2043.5E6	740.958	739.487
6) T	MCPD	5.738	5.378	303.7E6	153.5E6	81.313	78.556
7) T	MCPA	5.869	5.553	447.6E6	228.5E6	78.417	76.501
8) T	DICHLORPROP	6.210	5.823	1009.0E6	490.2E6	737.563	714.592
9) T	2,4-D	6.428	6.094	1038.2E6	582.3E6	689.060	715.721
10) T	Pentachlo...	6.679	6.453	15133.8E6	7489.6E6	743.868	736.316
11) T	2,4,5-TP ...	7.212	6.816	5991.2E6	3006.0E6	747.408	743.147
12) T	2,4,5-T	7.486	7.175	6329.0E6	2696.7E6	741.654	716.710
13) T	2,4-DB	8.012	7.659	787.3E6	385.9E6	696.747m	786.330
14) T	DINOSEB	9.089	7.965	4265.0E6	1902.0E6	752.918	748.735
15) T	Picloram	8.932	8.901	7963.1E6	4138.5E6	672.278	687.625
16) T	DCPA	9.362	8.855	7704.3E6	4038.7E6	761.920	762.773

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121919\
Data File : PS008106.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2019 17:25
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
12/20/2019 8:08:12 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 05:35:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121819.M
Quant Title : 8080.M
QLast Update : Wed Dec 18 11:09:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

