

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121919\
 Data File : PS008118.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2019 21:48
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 05:39:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121819.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 18 11:09:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.425	5.156	1201.2E6	572.1E6	801.785	779.445
Target Compounds							
1) T	Dalapon	2.010	1.722	1189.5E6	511.0E6	635.388	576.866
2) T	3,5-DICHL...	4.802	4.465	1423.2E6	679.7E6	726.560	720.857
3) T	4-Nitroph...	5.285	4.849	434.3E6	261.1E6	685.096	668.481
5) T	DICAMBA	5.573	5.279	4238.5E6	2080.3E6	765.753	752.800
6) T	MCPD	5.723	5.363	309.6E6	155.4E6	82.903	79.536
7) T	MCPA	5.854	5.538	457.9E6	229.8E6	80.226	76.946
8) T	DICHLORPROP	6.195	5.808	1043.1E6	508.5E6	762.469	741.316
9) T	2,4-D	6.412	6.079	1083.0E6	604.0E6	718.794	742.370
10) T	Pentachlo...	6.664	6.438	15538.0E6	7628.8E6	763.735	750.002
11) T	2,4,5-TP ...	7.196	6.801	6165.7E6	3071.2E6	769.189	759.268
12) T	2,4,5-T	7.469	7.160	6640.9E6	2799.8E6	778.195	744.125
13) T	2,4-DB	7.996	7.644	815.0E6	395.1E6	721.265	805.052
14) T	DINOSEB	9.074	7.949	4691.3E6	2115.2E6	828.187	832.666
15) T	Picloram	8.915	8.886	8471.3E6	4330.3E6	715.185	719.496
16) T	DCPA	9.348	8.840	7908.3E6	4094.3E6	782.091	773.278

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS121919\
Data File : PS008118.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2019 21:48
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 05:39:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121819.M
Quant Title : 8080.M
QLast Update : Wed Dec 18 11:09:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

