

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122319\
 Data File : PS008139.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2019 14:12
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 24 01:21:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121819.M
 Quant Title : 8080.M
 QLast Update : Wed Dec 18 11:09:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.431	5.162	1201.3E6	552.2E6	801.852	752.340
Target Compounds							
1) T	Dalapon	2.015	1.726	1372.3E6	573.4E6	733.064	647.363
2) T	3,5-DICHL...	4.808	4.471	1476.4E6	667.9E6	753.736	708.317
3) T	4-Nitroph...	5.288	4.854	490.8E6	277.1E6	774.239	709.212
5) T	DICAMBA	5.579	5.286	4232.6E6	1995.8E6	764.697	722.235
6) T	MCPD	5.730	5.370	310.2E6	144.5E6	83.055	73.972
7) T	MCPA	5.860	5.545	444.8E6	217.2E6	77.932	72.721
8) T	DICHLORPROP	6.201	5.815	1040.7E6	486.5E6	760.671	709.322
9) T	2,4-D	6.418	6.085	1127.5E6	590.7E6	748.362	725.976
10) T	Pentachlo...	6.671	6.445	15746.9E6	7470.4E6	774.003	734.427
11) T	2,4,5-TP ...	7.204	6.808	6209.6E6	2950.5E6	774.655	729.424
12) T	2,4,5-T	7.476	7.166	6703.1E6	2752.6E6	785.484	731.572
13) T	2,4-DB	8.002	7.650	879.9E6	373.7E6	778.752	761.458
14) T	DINOSEB	9.081	7.957	4419.7E6	1928.5E6	780.238	759.193
15) T	Picloram	8.919	8.892	9686.1E6	4707.2E6	817.742	782.118
16) T	DCPA	9.354	8.847	7794.2E6	3867.8E6	770.812	730.509

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS122319\
Data File : PS008139.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2019 14:12
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 24 01:21:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS121819.M
Quant Title : 8080.M
QLast Update : Wed Dec 18 11:09:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

