

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF010219\
 Data File : FF003126.D
 Signal(s) : FID2B.ch
 Acq On : 02 Jan 2019 15:31
 Operator : AJ\SJ
 Sample : PB116089BS
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 PB116089BS

Manual Integrations
 APPROVED

Sohil
 1/3/2019 8:58:21 AM

Integration File: autoint1.e
 Quant Time: Jan 02 23:51:42 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
 Quant Title :
 QLast Update : Tue Dec 04 16:40:56 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.630	2202533	16.965 ug/ml
Target Compounds			
1) N-OCTANE	2.260	2959136	19.129 ug/ml
2) N-DECANE	4.359	3038334	19.393 ug/ml
3) N-DODECANE	6.380	2982914	18.784 ug/ml
4) N-TETRADECANE	8.174	2901968	18.105 ug/ml
5) N-HEXADECANE	9.774	2814847	17.577 ug/ml
6) N-OCTADECANE	11.216	2740519	17.299 ug/ml
7) N-EICOSANE	12.528	2651835	17.194 ug/ml
8) N-DOCOSANE	13.728	2574649	17.218 ug/ml
10) N-TETRACOSANE	14.834	2580293	17.855 ug/ml
11) N-HEXACOSANE	15.858	2491795	17.852 ug/ml
12) N-OCTACOSANE	16.810	2528885	18.874 ug/ml
13) N-TRIACONTANE	17.700	2646086	20.651 ug/ml
14) N-DOTRIACONTANE	18.535	2674587	22.396 ug/ml
15) N-TETRATRIACONTANE	19.322	2575254	22.722 ug/ml
16) N-HEXATRIACONTANE	20.066	2390142	22.152 ug/ml
17) N-OCTATRIACONTANE	20.804	2127794	20.964 ug/mlm
18) N-TETRACONTANE	21.699	1996035	18.991 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF010219\
Data File : FF003126.D
Signal(s) : FID2B.ch
Acq On : 02 Jan 2019 15:31
Operator : AJ\SJ
Sample : PB116089BS
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
FID_F
ClientSampled :
PB116089BS

Manual Integrations
APPROVED

Sohil
1/3/2019 8:58:21 AM

Integration File: autoint1.e
Quant Time: Jan 02 23:51:42 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF120418.M
Quant Title :
QLast Update : Tue Dec 04 16:40:56 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

