

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF010421\
 Data File : FF008046.D
 Signal(s) : FID2B.ch
 Acq On : 04 Jan 2021 20:58
 Operator : DD\AJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 05 11:46:59 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF010421.M
 Quant Title :
 QLast Update : Tue Jan 05 11:37:03 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.676	1271501	10.445 ug/ml
Target Compounds			
1) N-OCTANE	2.299	1220061	10.067 ug/ml
2) N-DECANE	4.402	1233352	9.832 ug/ml
3) N-DODECANE	6.423	1285572	9.865 ug/ml
4) N-TETRADECANE	8.219	1359106	10.049 ug/ml
5) N-HEXADECANE	9.820	1404478	10.201 ug/ml
6) N-OCTADECANE	11.263	1452479	10.256 ug/ml
7) N-EICOSANE	12.574	1471267	10.353 ug/ml
8) N-DOCOSANE	13.774	1458051	10.414 ug/ml
10) N-TETRACOSANE	14.880	1465373	10.457 ug/ml
11) N-HEXACOSANE	15.902	1444188	10.509 ug/ml
12) N-OCTACOSANE	16.853	1420956	10.604 ug/ml
13) N-TRIACONTANE	17.742	1377909	10.758 ug/ml
14) N-DOTRIACONTANE	18.576	1239741	10.848 ug/ml
15) N-TETRATRIACONTANE	19.361	994122	10.765 ug/ml
16) N-HEXATRIACONTANE	20.102	752840	10.741 ug/mlm
17) N-OCTATRIACONTANE	20.846	612230	10.533 ug/ml
18) N-TETRACONTANE	21.761	610830	10.884 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF010421\
Data File : FF008046.D
Signal(s) : FID2B.ch
Acq On : 04 Jan 2021 20:58
Operator : DD\AJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 60 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 05 11:46:59 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF010421.M
Quant Title :
QLast Update : Tue Jan 05 11:37:03 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

