

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF011022\
Data File : FF009666.D
Signal(s) : FID2B.ch
Acq On : 10 Jan 2022 18:43
Operator : AJ\MA
Sample : RT MARKER
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
RT MARKER

Integration File: Sample.e
Quant Time: Jan 11 06:41:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF121521.M
Quant Title :
QLast Update : Thu Dec 16 03:33:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
Target Compounds			
3) N-DODECANE	6.249	4255483	29.426 ug/ml
12) N-OCTACOSANE	16.623	3634625	30.523 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF011022\
Data File : FF009666.D
Signal(s) : FID2B.ch
Acq On : 10 Jan 2022 18:43
Operator : AJ\MA
Sample : RT MARKER
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
RT MARKER

Integration File: Sample.e
Quant Time: Jan 11 06:41:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF121521.M
Quant Title :
QLast Update : Thu Dec 16 03:33:07 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE


R.T.: 6.249 min
Delta R.T.: -0.022 min
Response: 4255483
Conc: 29.43 ug/ml

Instrument :
FID_F
ClientSampleId :
RT MARKER



#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T. : 14.479 min
Response: 0
Conc: N.D.



#12 N-OCTACOSANE

R.T.: 16.623 min
Delta R.T.: -0.023 min
Response: 3634625
Conc: 30.52 ug/ml