

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF011320\
Data File : FF006180.D
Signal(s) : FID2B.ch
Acq On : 13 Jan 2020 17:40
Operator : DD\AJ
Sample : PB126032BL
Misc :
ALS Vial : 61 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 14 07:54:45 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011020.M
Quant Title :
QLast Update : Fri Jan 10 12:35:13 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.623	2857087	26.367 ug/ml
Target Compounds			
4) N-TETRADECANE	8.250f	185458	1.469 ug/ml
11) N-HEXACOSANE	15.841	82096	0.668 ug/ml
12) N-OCTACOSANE	16.786	182447	1.501 ug/ml
13) N-TRIACONTANE	17.671	206953	1.726 ug/ml
14) N-DOTRIACONTANE	18.496	447395	3.871 ug/ml
15) N-TETRATRIACONTANE	19.281	124757	1.107 ug/ml
16) N-HEXATRIACONTANE	20.050	281663	2.634 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF011320\
Data File : FF006180.D
Signal(s) : FID2B.ch
Acq On : 13 Jan 2020 17:40
Operator : DD\AJ
Sample : PB126032BL
Misc :
ALS Vial : 61 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 14 07:54:45 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011020.M
Quant Title :
QLast Update : Fri Jan 10 12:35:13 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#4 N-TETRADECANE

R.T.: 8.250 min
Delta R.T.: 0.060 min
Response: 185458
Conc: 1.47 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.623 min
Delta R.T.: 0.003 min
Response: 2857087
Conc: 26.37 ug/ml




#11 N-HEXACOSANE

R.T.: 15.841 min
Delta R.T.: 0.000 min
Response: 82096
Conc: 0.67 ug/ml




#12 N-OCTACOSANE

R.T.: 16.786 min
Delta R.T.: 0.000 min
Response: 182447
Conc: 1.50 ug/ml




#13 N-TRIACONTANE

R.T.: 17.671 min
Delta R.T.: 0.000 min
Response: 206953
Conc: 1.73 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.496 min
Delta R.T.: -0.005 min
Response: 447395
Conc: 3.87 ug/ml



#15 N-TETRATRIACONTANE

R.T.: 19.281 min
Delta R.T.: -0.002 min
Response: 124757
Conc: 1.11 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.050 min
Delta R.T.: 0.029 min
Response: 281663
Conc: 2.63 ug/ml