

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF011419\
Data File : FF003237.D
Signal(s) : FID2B.ch
Acq On : 14 Jan 2019 11:56
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 14 21:32:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF010419.M
Quant Title :
QLast Update : Fri Jan 04 16:10:51 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

Target Compounds

3)	N-DODECANE	6.368	2875316	18.999 ug/ml
12)	N-OCTACOSANE	16.789	2562394	17.606 ug/ml
13)	N-TRIACONTANE	17.676	94888	0.664 ug/ml
14)	N-DOTRIACONTANE	18.514	116546	0.875 ug/ml
16)	N-HEXATRIACONTANE	20.080	188025	1.667 ug/ml
18)	N-TETRACONTANE	21.698	297655	2.953 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF011419\
Data File : FF003237.D
Signal(s) : FID2B.ch
Acq On : 14 Jan 2019 11:56
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 14 21:32:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF010419.M
Quant Title :
QLast Update : Fri Jan 04 16:10:51 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE

R.T.: 6.368 min
Delta R.T.: -0.012 min
Response: 2875316
Conc: 19.00 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T. : 14.630 min
Response: 0
Conc: N.D.




#12 N-OCTACOSANE

R.T.: 16.789 min
Delta R.T.: -0.023 min
Response: 2562394
Conc: 17.61 ug/ml




#13 N-TRIACONTANE

R.T.: 17.676 min
Delta R.T.: -0.024 min
Response: 94888
Conc: 0.66 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.514 min
Delta R.T.: -0.021 min
Response: 116546
Conc: 0.87 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.080 min
Delta R.T.: 0.014 min
Response: 188025
Conc: 1.67 ug/ml



#18 N-TETRACONTANE

R.T.: 21.698 min
Delta R.T.: -0.004 min
Response: 297655
Conc: 2.95 ug/ml