

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF011819\
 Data File : FF003291.D
 Signal(s) : FID2B.ch
 Acq On : 18 Jan 2019 09:10
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :

Manual Integrations
 APPROVED

Sohil
 1/22/2019 11:42:22 AM

Integration File: Sample.e
 Quant Time: Jan 18 20:52:43 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF010419.M
 Quant Title :
 QLast Update : Tue Jan 15 08:58:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.612	6247385	43.662 ug/ml
Target Compounds			
1) N-OCTANE	2.256	8390734	51.952 ug/ml
2) N-DECANE	4.350	8706640	53.837 ug/ml
3) N-DODECANE	6.368	8528376	52.574 ug/ml
4) N-TETRADECANE	8.159	8282625	50.725 ug/ml
5) N-HEXADECANE	9.758	8037687	48.950 ug/ml
6) N-OCTADECANE	11.200	7798453	47.160 ug/ml
7) N-EICOSANE	12.511	7512313	45.753 ug/ml
8) N-DOCOSANE	13.711	7282647	44.819 ug/ml
10) N-TETRACOSANE	14.816	7129299	44.452 ug/ml
11) N-HEXACOSANE	15.840	7072113	44.712 ug/ml
12) N-OCTACOSANE	16.793	7245279	45.999 ug/ml
13) N-TRIACONTANE	17.682	7565443	48.188 ug/ml
14) N-DOTRIACONTANE	18.519	7594088	52.222 ug/ml
15) N-TETRATRIACONTANE	19.305	7431514	56.211 ug/mlm
16) N-HEXATRIACONTANE	20.049	7098377	60.139 ug/mlm
17) N-OCTATRIACONTANE	20.782	6516499	62.364 ug/ml
18) N-TETRACONTANE	21.682	6522558	61.247 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF011819\
Data File : FF003291.D
Signal(s) : FID2B.ch
Acq On : 18 Jan 2019 09:10
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampled :

Manual Integrations
APPROVED

Sohil
1/22/2019 11:42:22 AM

Integration File: Sample.e
Quant Time: Jan 18 20:52:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF010419.M
Quant Title :
QLast Update : Tue Jan 15 08:58:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rx1-1ms
Signal Info : 20mx0.18mmx0.18um

