

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012022\
 Data File : FF009719.D
 Signal(s) : FID2B.ch
 Acq On : 20 Jan 2022 10:16
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/21/2022
 Supervised By :Ankita Jodhani 01/21/2022

Integration File: autoint1.e
 Quant Time: Jan 20 23:46:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011722.M
 Quant Title :
 QLast Update : Tue Jan 18 02:53:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.449	6695966	51.240 ug/ml
Target Compounds			
1) N-OCTANE	2.173	7445376	53.858 ug/ml
2) N-DECANE	4.237	7762180	54.884 ug/ml
3) N-DODECANE	6.245	7837744	54.771 ug/ml
4) N-TETRADECANE	8.029	7913419	53.852 ug/ml
5) N-HEXADECANE	9.621	7921754	52.955 ug/ml
6) N-OCTADECANE	11.057	8098630	52.531 ug/ml
7) N-EICOSANE	12.361	7832126	51.813 ug/ml
8) N-DOCOSANE	13.553	7679185	51.636 ug/ml
10) N-TETRACOSANE	14.653	7508902	50.633 ug/ml
11) N-HEXACOSANE	15.669	7286375	49.836 ug/ml
12) N-OCTACOSANE	16.613	6975884	48.055 ug/ml
13) N-TRIACONTANE	17.498	6637217	46.387 ug/ml
14) N-DOTRIACONTANE	18.327	6321128	46.451 ug/ml
15) N-TETRATRIACONTANE	19.107	5875392	49.467 ug/ml
16) N-HEXATRIACONTANE	19.843	5297265	51.378 ug/ml
17) N-OCTATRIACONTANE	20.547	4639176	52.025 ug/mlm
18) N-TETRACONTANE	21.366	4716313	51.920 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012022\
Data File : FF009719.D
Signal(s) : FID2B.ch
Acq On : 20 Jan 2022 10:16
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/21/2022
Supervised By :Ankita Jodhani 01/21/2022

Integration File: autoint1.e
Quant Time: Jan 20 23:46:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011722.M
Quant Title :
QLast Update : Tue Jan 18 02:53:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

