

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012121\
 Data File : FF008149.D
 Signal(s) : FID2B.ch
 Acq On : 21 Jan 2021 22:09
 Operator : DD\AJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 22 00:25:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF012121.M
 Quant Title :
 QLast Update : Fri Jan 22 00:23:11 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.671	1291429	10.559 ug/ml
Target Compounds			
1) N-OCTANE	2.299	1366779	10.069 ug/ml
2) N-DECANE	4.398	1387713	9.926 ug/ml
3) N-DODECANE	6.419	1431110	10.039 ug/ml
4) N-TETRADECANE	8.214	1495168	10.220 ug/ml
5) N-HEXADECANE	9.816	1502924	10.329 ug/ml
6) N-OCTADECANE	11.259	1524077	10.413 ug/ml
7) N-EICOSANE	12.570	1513269	10.511 ug/ml
8) N-DOCOSANE	13.770	1491236	10.564 ug/ml
10) N-TETRACOSANE	14.875	1465078	10.586 ug/ml
11) N-HEXACOSANE	15.898	1412565	10.571 ug/ml
12) N-OCTACOSANE	16.850	1373280	10.575 ug/ml
13) N-TRIACONTANE	17.739	1331679	10.644 ug/ml
14) N-DOTRIACONTANE	18.573	1339435	11.170 ug/ml
15) N-TETRATRIACONTANE	19.357	1218752	10.548 ug/ml
16) N-HEXATRIACONTANE	20.099	1138682	10.453 ug/ml
17) N-OCTATRIACONTANE	20.840	1066130	10.310 ug/ml
18) N-TETRACONTANE	21.757	1206986	10.846 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012121\
Data File : FF008149.D
Signal(s) : FID2B.ch
Acq On : 21 Jan 2021 22:09
Operator : DD\AJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 61 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 22 00:25:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF012121.M
Quant Title :
QLast Update : Fri Jan 22 00:23:11 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

