

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012419\
 Data File : FF003324.D
 Signal(s) : FID2B.ch
 Acq On : 24 Jan 2019 10:32
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 25 03:15:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF010419.M
 Quant Title :
 QLast Update : Tue Jan 15 08:58:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.613	6491717	45.369 ug/ml
Target Compounds			
2) N-DECANE	4.347	8880876	54.914 ug/ml
3) N-DODECANE	6.366	8731909	53.829 ug/ml
4) N-TETRADECANE	8.158	8502494	52.071 ug/ml
5) N-HEXADECANE	9.758	8282026	50.438 ug/ml
6) N-OCTADECANE	11.200	8031261	48.568 ug/ml
7) N-EICOSANE	12.511	7698401	46.886 ug/ml
8) N-DOCOSANE	13.712	7435273	45.758 ug/ml
10) N-TETRACOSANE	14.818	7247354	45.188 ug/ml
11) N-HEXACOSANE	15.843	7155553	45.240 ug/ml
12) N-OCTACOSANE	16.796	7163248	45.478 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012419\
Data File : FF003324.D
Signal(s) : FID2B.ch
Acq On : 24 Jan 2019 10:32
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 25 03:15:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF010419.M
Quant Title :
QLast Update : Tue Jan 15 08:58:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

