

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012420\
 Data File : FF006299.D
 Signal(s) : FID2B.ch
 Acq On : 25 Jan 2020 00:55
 Operator : DD\AJ
 Sample : L1161-09
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 MDL-WATER-03-QT1-2020

Integration File: autoint1.e
 Quant Time: Jan 25 04:12:12 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011020.M
 Quant Title :
 QLast Update : Fri Jan 10 12:35:13 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.623	1745667	16.110 ug/ml
Target Compounds			
2) N-DECANE	4.383	166054	1.332 ug/ml
3) N-DODECANE	6.399	151172	1.197 ug/ml
4) N-TETRADECANE	8.190	122484	0.970 ug/ml
5) N-HEXADECANE	9.787	155665	1.218 ug/ml
6) N-OCTADECANE	11.224	152934	1.206 ug/ml
7) N-EICOSANE	12.530	154624	1.210 ug/ml
8) N-DOCOSANE	13.724	152646	1.211 ug/ml
10) N-TETRACOSANE	14.824	152672	1.226 ug/ml
11) N-HEXACOSANE	15.842	160964	1.309 ug/ml
12) N-OCTACOSANE	16.788	191332	1.574 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012420\
Data File : FF006299.D
Signal(s) : FID2B.ch
Acq On : 25 Jan 2020 00:55
Operator : DD\AJ
Sample : L1161-09
Misc :
ALS Vial : 72 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
MDL-WATER-03-QT1-2020

Integration File: autoint1.e
Quant Time: Jan 25 04:12:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011020.M
Quant Title :
QLast Update : Fri Jan 10 12:35:13 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

