

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012519\
 Data File : FF003335.D
 Signal(s) : FID2B.ch
 Acq On : 25 Jan 2019 09:10
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 27 23:42:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF010419.M
 Quant Title :
 QLast Update : Tue Jan 15 08:58:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.616	6748002	47.160 ug/ml
Target Compounds			
1) N-OCTANE	2.253	8880537	54.985 ug/ml
2) N-DECANE	4.349	9204188	56.913 ug/ml
3) N-DODECANE	6.368	9028718	55.658 ug/ml
4) N-TETRADECANE	8.160	8787701	53.818 ug/ml
5) N-HEXADECANE	9.759	8560864	52.136 ug/ml
6) N-OCTADECANE	11.202	8298598	50.184 ug/ml
7) N-EICOSANE	12.513	7973202	48.560 ug/ml
8) N-DOCOSANE	13.714	7716991	47.492 ug/ml
10) N-TETRACOSANE	14.820	7543772	47.036 ug/ml
11) N-HEXACOSANE	15.844	7476842	47.271 ug/ml
12) N-OCTACOSANE	16.798	7621645	48.389 ug/ml
13) N-TRIACONTANE	17.690	8062667	51.355 ug/ml
14) N-DOTRIACONTANE	18.525	8179465	56.248 ug/ml
15) N-TETRATRIACONTANE	19.313	7970305	60.287 ug/mlm
16) N-HEXATRIACONTANE	20.058	7542530	63.902 ug/mlm
17) N-OCTATRIACONTANE	20.797	6787153	64.955 ug/ml
18) N-TETRACONTANE	21.696	6884651	64.648 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012519\
Data File : FF003335.D
Signal(s) : FID2B.ch
Acq On : 25 Jan 2019 09:10
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jan 27 23:42:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF010419.M
Quant Title :
QLast Update : Tue Jan 15 08:58:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

