

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012820\
 Data File : FF006344.D
 Signal(s) : FID2B.ch
 Acq On : 28 Jan 2020 21:40
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Sohil
 1/30/2020 3:19:50 PM

Integration File: autoint1.e
 Quant Time: Jan 29 00:34:18 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011020.M
 Quant Title :
 QLast Update : Fri Jan 10 12:35:13 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.628	5711687	52.711 ug/ml
Target Compounds			
1) N-OCTANE	2.289	7603778	61.901 ug/ml
2) N-DECANE	4.385	7866927	63.118 ug/ml
3) N-DODECANE	6.403	7881189	62.409 ug/ml
4) N-TETRADECANE	8.195	7673130	60.790 ug/ml
5) N-HEXADECANE	9.792	7499135	58.679 ug/ml
6) N-OCTADECANE	11.230	7162276	56.465 ug/ml
7) N-EICOSANE	12.537	6962921	54.480 ug/ml
8) N-DOCOSANE	13.731	6699500	53.132 ug/ml
10) N-TETRACOSANE	14.832	6550547	52.585 ug/ml
11) N-HEXACOSANE	15.850	6445638	52.431 ug/ml
12) N-OCTACOSANE	16.797	6453073	53.084 ug/ml
13) N-TRIACONTANE	17.682	6427464	53.607 ug/ml
14) N-DOTRIACONTANE	18.511	6109074	52.863 ug/ml
15) N-TETRATRIACONTANE	19.293	5705461	50.643 ug/mlm
16) N-HEXATRIACONTANE	20.028	5215221	48.776 ug/mlm
17) N-OCTATRIACONTANE	20.754	4736092	51.192 ug/ml
18) N-TETRACONTANE	21.640	4747766	47.150 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF012820\
Data File : FF006344.D
Signal(s) : FID2B.ch
Acq On : 28 Jan 2020 21:40
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Sohil
1/30/2020 3:19:50 PM

Integration File: autoint1.e
Quant Time: Jan 29 00:34:18 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011020.M
Quant Title :
QLast Update : Fri Jan 10 12:35:13 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

