

Method Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\
 Method File : FF020119.M
 Title :
 Last Update : Fri Feb 01 22:57:49 2019
 Response Via : Initial Calibration

Calibration Files

50 =FF003350.D 100 =FF003349.D 20 =FF003351.D
 10 =FF003352.D 5 =FF003353.D

Compound		50	100	20	10	5	Avg	%RSD
1)	N-OCTANE	1.263	1.276	1.377	1.321	1.438	1.335 E5	5.45
2)	N-DECANE	1.278	1.292	1.387	1.320	1.415	1.338 E5	4.49
3)	N-DODECANE	1.289	1.302	1.399	1.330	1.421	1.348 E5	4.35
4)	N-TETRADECANE	1.296	1.308	1.404	1.335	1.424	1.354 E5	4.24
5)	N-HEXADECANE	1.307	1.318	1.405	1.345	1.439	1.363 E5	4.17
6)	N-OCTADECANE	1.314	1.322	1.420	1.358	1.466	1.376 E5	4.77
7)	N-EICOSANE	1.303	1.309	1.410	1.361	1.492	1.375 E5	5.70
8)	N-DOCOSANE	1.293	1.298	1.397	1.353	1.500	1.368 E5	6.23
9) S	TETRACOSANE-d50 (S...	1.140	1.146	1.229	1.195	1.326	1.207 E5	6.29
10)	N-TETRACOSANE	1.281	1.286	1.385	1.345	1.503	1.360 E5	6.67
11)	N-HEXACOSANE	1.277	1.281	1.379	1.342	1.503	1.356 E5	6.81
12)	N-OCTACOSANE	1.279	1.277	1.384	1.362	1.543	1.369 E5	7.93
13)	N-TRIACONTANE	1.291	1.276	1.416	1.428	1.649	1.412 E5	10.60
14)	N-DOTRIACONTANE	1.244	1.225	1.369	1.413	1.648	1.380 E5	12.31
15)	N-TETRATRIACONTANE	1.187	1.179	1.280	1.262	1.435	1.268 E5	8.14
16)	N-HEXATRIACONTANE	1.129	1.116	1.217	1.125	1.258	1.169 E5	5.49
17)	N-OCTATRIACONTANE	1.014	1.010	1.097	1.077	1.206	1.081 E5	7.39
18)	N-TETRACONTANE	1.000	0.993	1.073	1.213	1.135	1.083 E5	8.60

(#) = Out of Range