

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF020122\
Data File : FF009821.D
Signal(s) : FID2B.ch
Acq On : 01 Feb 2022 10:58
Operator : AJ\MA
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
RT MARKER

Integration File: calibration.e
Quant Time: Feb 02 01:59:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011722.M
Quant Title :
QLast Update : Tue Jan 18 02:53:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
Target Compounds			
3) N-DODECANE	6.229	4245174	29.666 ug/ml
12) N-OCTACOSANE	16.597	3739311	25.759 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF020122\
Data File : FF009821.D
Signal(s) : FID2B.ch
Acq On : 01 Feb 2022 10:58
Operator : AJ\MA
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
RT MARKER

Integration File: calibration.e
Quant Time: Feb 02 01:59:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011722.M
Quant Title :
QLast Update : Tue Jan 18 02:53:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE


R.T.: 6.229 min
Delta R.T.: -0.020 min
Response: 4245174
Conc: 29.67 ug/ml

Instrument :
FID_F
ClientSampleId :
RT MARKER



#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T. : 14.457 min
Response: 0
Conc: N.D.



#12 N-OCTACOSANE

R.T.: 16.597 min
Delta R.T.: -0.027 min
Response: 3739311
Conc: 25.76 ug/ml