

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF020124\
 Data File : FF013941.D
 Signal(s) : FID2B.ch
 Acq On : 01 Feb 2024 12:25
 Operator : YP\AJ
 Sample : 20 TRPH STD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 01 14:09:14 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020124.M
 Quant Title :
 QLast Update : Thu Feb 01 14:05:46 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.953	2089056	20.037 ug/ml
Target Compounds			
1) N-OCTANE	1.991	2226702	20.657 ug/ml
2) N-DECANE	4.507	2143764	20.289 ug/ml
3) N-DODECANE	6.670	2214120	20.142 ug/ml
4) N-TETRADECANE	8.495	2182459	20.074 ug/ml
5) N-HEXADECANE	10.101	2254806	20.087 ug/ml
6) N-OCTADECANE	11.544	2343119	20.020 ug/ml
7) N-EICOSANE	12.853	2312100	19.992 ug/ml
8) N-DOCOSANE	14.053	2297404	20.016 ug/ml
10) N-TETRACOSANE	15.157	2273434	19.975 ug/ml
11) N-HEXACOSANE	16.178	2201465	19.921 ug/ml
12) N-OCTACOSANE	17.128	2129864	19.907 ug/ml
13) N-TRIACONTANE	18.018	2086593	20.098 ug/ml
14) N-DOTRIACONTANE	18.850	1993437	20.390 ug/ml
15) N-TETRATRIACONTANE	19.633	1602972	20.475 ug/ml
16) N-HEXATRIACONTANE	20.374	1227332	20.357 ug/ml
17) N-OCTATRIACONTANE	21.101	969587	19.618 ug/ml
18) N-TETRACONTANE	21.989	924053	21.943 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF020124\
Data File : FF013941.D
Signal(s) : FID2B.ch
Acq On : 01 Feb 2024 12:25
Operator : YP\AJ
Sample : 20 TRPH STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 01 14:09:14 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020124.M
Quant Title :
QLast Update : Thu Feb 01 14:05:46 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

