

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF020322\
 Data File : FF009857.D
 Signal(s) : FID2B.ch
 Acq On : 04 Feb 2022 01:09
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/04/2022
 Supervised By :Ankita Jodhani 02/04/2022

Integration File: calibration.e
 Quant Time: Feb 04 01:28:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011722.M
 Quant Title :
 QLast Update : Tue Jan 18 02:53:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.433	5587465	42.757 ug/mlm
Target Compounds			
1) N-OCTANE	2.158	7059381	51.066 ug/ml
2) N-DECANE	4.222	7077175	50.040 ug/ml
3) N-DODECANE	6.228	7077526	49.459 ug/ml
4) N-TETRADECANE	8.014	7085253	48.216 ug/ml
5) N-HEXADECANE	9.606	6994665	46.758 ug/ml
6) N-OCTADECANE	11.041	7045487	45.700 ug/ml
7) N-EICOSANE	12.345	6691977	44.270 ug/mlm
8) N-DOCOSANE	13.537	6481385	43.582 ug/ml
10) N-TETRACOSANE	14.637	6330709	42.689 ug/ml
11) N-HEXACOSANE	15.654	6163988	42.159 ug/ml
12) N-OCTACOSANE	16.600	6168691	42.494 ug/ml
13) N-TRIACONTANE	17.482	6006726	41.980 ug/mlm
14) N-DOTRIACONTANE	18.311	6039192	44.379 ug/mlm
15) N-TETRATRIACONTANE	19.092	5713965	48.108 ug/mlm
16) N-HEXATRIACONTANE	19.828	5248124	50.902 ug/mlm
17) N-OCTATRIACONTANE	20.532	4508453	50.559 ug/mlm
18) N-TETRACONTANE	21.349	4897180	53.911 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF020322\
Data File : FF009857.D
Signal(s) : FID2B.ch
Acq On : 04 Feb 2022 01:09
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 02/04/2022
Supervised By : Ankita Jodhani 02/04/2022

Integration File: calibration.e
Quant Time: Feb 04 01:28:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011722.M
Quant Title :
QLast Update : Tue Jan 18 02:53:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

