

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF020422\
 Data File : FF009880.D
 Signal(s) : FID2B.ch
 Acq On : 04 Feb 2022 23:21
 Operator : AJ\MA
 Sample : N1393-03MSD
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 SP25-03-20220203MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/17/2022
 Supervised By :Ankita Jodhani 02/17/2022

Integration File: calibration.e
 Quant Time: Feb 17 02:20:55 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011722.M
 Quant Title :
 QLast Update : Tue Jan 18 02:53:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.426	2247520	17.199 ug/ml
Target Compounds			
1) N-OCTANE	2.158	1473931	10.662 ug/ml
2) N-DECANE	4.217	2127372	15.042 ug/ml
3) N-DODECANE	6.221	2971202	20.763 ug/ml
4) N-TETRADECANE	8.006	2906727	19.781 ug/ml
5) N-HEXADECANE	9.598	2621875	17.527 ug/mlm
6) N-OCTADECANE	11.034	4089292	26.525 ug/mlm
7) N-EICOSANE	12.338	2423545	16.033 ug/mlm
8) N-DOCOSANE	13.530	2344554	15.765 ug/ml
10) N-TETRACOSANE	14.628	2296429	15.485 ug/ml
11) N-HEXACOSANE	15.644	2243198	15.343 ug/ml
12) N-OCTACOSANE	16.590	2235331	15.399 ug/mlm
13) N-TRIACONTANE	17.473	2274357	15.895 ug/ml
14) N-DOTRIACONTANE	18.302	2255946	16.578 ug/ml
15) N-TETRATRIACONTANE	19.083	1994542	16.793 ug/ml
16) N-HEXATRIACONTANE	19.819	1813987	17.594 ug/mlm
17) N-OCTATRIACONTANE	20.521	1634767	18.333 ug/mlm
18) N-TETRACONTANE	21.331	1946333	21.427 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF020422\
Data File : FF009880.D
Signal(s) : FID2B.ch
Acq On : 04 Feb 2022 23:21
Operator : AJ\MA
Sample : N1393-03MSD
Misc :
ALS Vial : 66 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
SP25-03-20220203MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/17/2022
Supervised By :Ankita Jodhani 02/17/2022

Integration File: calibration.e
Quant Time: Feb 17 02:20:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011722.M
Quant Title :
QLast Update : Tue Jan 18 02:53:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

