

rteres

Instrument :

FID_F

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 02/07/2022
Supervised By :Ankita Jodhani 02/07/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF02042
Data File : FF009860.D
Signal(s) : FID2B.ch
Acq On : 04 Feb 2022 09:56
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF011722.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.157	2.116	2.250	PH	984634	7132799	97.21%	6.181%
2	4.220	4.190	4.298	BH	806377	7336975	99.99%	6.357%
3	6.227	6.183	6.338	PH	757602	7337764	100.00%	6.358%
4	8.013	7.959	8.070	BH	716111	7321542	99.78%	6.344%
5	9.605	9.533	9.700	BHA	667817	7242917	98.71%	6.276%
6	11.040	11.005	11.100	HHA	643312	7300645	99.49%	6.326%
7	12.344	12.306	12.422	M	593880	6928524	94.42%	6.004%
8	13.538	13.502	13.584	HH	565491	6689400	91.16%	5.796%
9	14.433	14.388	14.464	PH	433226	5764526	78.56%	4.995%
10	14.637	14.555	14.696	PH	521707	6509479	88.71%	5.640%
11	15.654	15.572	15.693	PH	494073	6324185	86.19%	5.480%
12	16.600	16.535	16.665	HH	468332	6308920	85.98%	5.467%
13	17.484	17.399	17.552	HH	445820	6326742	86.22%	5.482%
14	18.312	18.259	18.378	M	435022	6162104	83.98%	5.339%
15	19.092	19.047	19.166	M	405188	5875073	80.07%	5.091%
16	19.830	19.782	19.913	M	346728	5454765	74.34%	4.727%
17	20.532	20.490	20.606	M	287684	4673857	63.70%	4.050%
18	21.345	21.282	21.477	M	211613	4717428	64.29%	4.088%
Sum of corrected areas:						115407645		

FF011722.M Mon Feb 07 11:57:39 2022