

Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID_F\Data\FF020719\
Data File : FF003417.D
Signal(s) : FID2B.ch
Acq On : 07 Feb 2019 17:26
Operator : AJ\SJ
Sample : K1392-03
Misc :
ALS Vial : 60 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 08 08:04:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
Quant Title :
QLast Update : Fri Feb 01 22:57:49 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.618	2001559	16.579 ug/ml
Target Compounds			
12) N-OCTACOSANE	16.800	176232	1.287 ug/ml
13) N-TRIACONTANE	17.690	218214	1.546 ug/ml
14) N-DOTRIACONTANE	18.527	176486	1.279 ug/ml
16) N-HEXATRIACONTANE	20.102f	204292	1.748 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID_F\Data\FF020719\
Data File : FF003417.D
Signal(s) : FID2B.ch
Acq On : 07 Feb 2019 17:26
Operator : AJ\SJ
Sample : K1392-03
Misc :
ALS Vial : 60 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 08 08:04:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
Quant Title :
QLast Update : Fri Feb 01 22:57:49 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.618 min
Delta R.T.: 0.008 min
Response: 2001559
Conc: 16.58 ug/ml




#12 N-OCTACOSANE

R.T.: 16.800 min
Delta R.T.: 0.008 min
Response: 176232
Conc: 1.29 ug/ml




#13 N-TRIACONTANE

R.T.: 17.690 min
Delta R.T.: 0.007 min
Response: 218214
Conc: 1.55 ug/ml



#14 N-DOTRIACONTANE

R.T.: 18.527 min
Delta R.T.: 0.007 min
Response: 176486
Conc: 1.28 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.102 min
Delta R.T.: 0.051 min
Response: 204292
Conc: 1.75 ug/ml