

Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID_F\Data\FF020719\
Data File : FF003421.D
Signal(s) : FID2B.ch
Acq On : 07 Feb 2019 19:56
Operator : AJ\SJ
Sample : K1392-05
Misc :
ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 08 08:04:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
Quant Title :
QLast Update : Fri Feb 01 22:57:49 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.619	2106725	17.450 ug/ml
Target Compounds			
12) N-OCTACOSANE	16.799	150548	1.100 ug/ml
13) N-TRIACONTANE	17.691	182956	1.296 ug/ml
14) N-DOTRIACONTANE	18.526	161955	1.174 ug/ml
16) N-HEXATRIACONTANE	20.106f	283394	2.424 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID_F\Data\FF020719\
Data File : FF003421.D
Signal(s) : FID2B.ch
Acq On : 07 Feb 2019 19:56
Operator : AJ\SJ
Sample : K1392-05
Misc :
ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 08 08:04:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
Quant Title :
QLast Update : Fri Feb 01 22:57:49 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.619 min
Delta R.T.: 0.009 min
Response: 2106725
Conc: 17.45 ug/ml




#12 N-OCTACOSANE

R.T.: 16.799 min
Delta R.T.: 0.007 min
Response: 150548
Conc: 1.10 ug/ml




#13 N-TRIACONTANE

R.T.: 17.691 min
Delta R.T.: 0.008 min
Response: 182956
Conc: 1.30 ug/ml



#14 N-DOTRIACONTANE

R.T.: 18.526 min
Delta R.T.: 0.007 min
Response: 161955
Conc: 1.17 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.106 min
Delta R.T.: 0.055 min
Response: 283394
Conc: 2.42 ug/ml