

Data Path : P:\HPCHEM1\FID_F\Data\FF020818\
 Data File : FF000073.D
 Signal(s) : FID2B.ch
 Acq On : 08 Feb 2018 14:27
 Operator : UA\AJ
 Sample : WATER MDL 05
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 08 14:51:09 2018
 Quant Method : P:\HPCHEM1\FID_F\Method\FF020118.M
 Quant Title :
 QLast Update : Fri Feb 02 10:05:00 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.632	1984944	17.428 ug/ml
Target Compounds			
1) N-OCTANE	2.287	203522	1.344 ug/ml
2) N-DECANE	4.378	225877	1.459 ug/ml
3) N-DODECANE	6.394	226827	1.469 ug/ml
4) N-TETRADECANE	8.186	225998	1.480 ug/ml
5) N-HEXADECANE	9.785	224272	1.525 ug/ml
6) N-OCTADECANE	11.225	216022	1.485 ug/ml
7) N-EICOSANE	12.534	201635	1.476 ug/ml
8) N-DOCOSANE	13.732	186631	1.432 ug/ml
10) N-TETRACOSANE	14.834	168303	1.328 ug/ml
11) N-HEXACOSANE	15.854	159205	1.313 ug/mlm
12) N-OCTACOSANE	16.803	174349	1.501 ug/mlm
13) N-TRIACONTANE	17.691	207279	1.893 ug/ml
14) N-DOTRIACONTANE	18.524	223319	2.265 ug/ml
15) N-TETRATRIACONTANE	19.310	201547	2.440 ug/ml
16) N-HEXATRIACONTANE	20.050	171571	2.618 ug/mlm
17) N-OCTATRIACONTANE	20.778	134089	2.219 ug/mlm
18) N-TETRACONTANE	21.660	100153	1.755 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : P:\HPCHEM1\FID_F\Data\FF020818\
Data File : FF000073.D
Signal(s) : FID2B.ch
Acq On : 08 Feb 2018 14:27
Operator : UA\AJ
Sample : WATER MDL 05
Misc :
ALS Vial : 61 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 08 14:51:09 2018
Quant Method : P:\HPCHEM1\FID_F\Method\FF020118.M
Quant Title :
QLast Update : Fri Feb 02 10:05:00 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

