

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF020819\
Data File : FF003430.D
Signal(s) : FID2B.ch
Acq On : 08 Feb 2019 11:13
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 08 21:45:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
Quant Title :
QLast Update : Fri Feb 01 22:57:49 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

Target Compounds			
3) N-DODECANE	6.363	3689869	27.367 ug/ml
12) N-OCTACOSANE	16.801	3645491	26.626 ug/ml
13) N-TRIACONTANE	17.689	522233	3.699 ug/ml
14) N-DOTRIACONTANE	18.527	747626	5.417 ug/ml
15) N-TETRATRIACONTANE	19.314	467832	3.688 ug/ml
16) N-HEXATRIACONTANE	20.059	287297	2.458 ug/ml
17) N-OCTATRIACONTANE	20.798	168166	1.556 ug/ml
18) N-TETRACONTANE	21.727	978917	9.041 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF020819\
Data File : FF003430.D
Signal(s) : FID2B.ch
Acq On : 08 Feb 2019 11:13
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 08 21:45:45 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
Quant Title :
QLast Update : Fri Feb 01 22:57:49 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE

R.T.: 6.363 min
Delta R.T.: 0.002 min
Response: 3689869
Conc: 27.37 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T.: 14.610 min
Response: 0
Conc: N.D.




#12 N-OCTACOSANE

R.T.: 16.801 min
Delta R.T.: 0.010 min
Response: 3645491
Conc: 26.63 ug/ml




#13 N-TRIACONTANE

R.T.: 17.689 min
Delta R.T.: 0.006 min
Response: 522233
Conc: 3.70 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.527 min
Delta R.T.: 0.007 min
Response: 747626
Conc: 5.42 ug/ml




#15 N-TETRATRIACONTANE

R.T.: 19.314 min
Delta R.T.: 0.007 min
Response: 467832
Conc: 3.69 ug/ml




#16 N-HEXATRIACONTANE

R.T.: 20.059 min
Delta R.T.: 0.009 min
Response: 287297
Conc: 2.46 ug/ml



#17 N-OCTATRIACONTANE

R.T.: 20.798 min
Delta R.T.: 0.011 min
Response: 168166
Conc: 1.56 ug/ml



#18 N-TETRACONTANE

R.T.: 21.727 min
Delta R.T.: 0.043 min
Response: 978917
Conc: 9.04 ug/ml