

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF021423\
Data File : FF012066.D
Signal(s) : FID2B.ch
Acq On : 14 Feb 2023 09:30
Operator : YP\AJ
Sample : MECL2
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :

Integration File: autoint1.e
Quant Time: Feb 14 16:44:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF012723.M
Quant Title :
QLast Update : Fri Jan 27 14:21:25 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
Target Compounds				
14)	N-DOTRIACONTANE	18.948	28468	0.219 ug/ml
15)	N-TETRATRIACONTANE	19.738	80363	0.722 ug/ml
16)	N-HEXATRIACONTANE	20.474	122933	1.197 ug/ml
18)	N-TETRACONTANE	22.179	87592	0.845 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF021423\
Data File : FF012066.D
Signal(s) : FID2B.ch
Acq On : 14 Feb 2023 09:30
Operator : YP\AJ
Sample : MECL2
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :

Integration File: autoint1.e
Quant Time: Feb 14 16:44:07 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF012723.M
Quant Title :
QLast Update : Fri Jan 27 14:21:25 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)


R.T.: 0.000 min
Exp R.T.: 15.053 min
Response: 0
Conc: N.D.

Instrument :
FID_F
ClientSampleId :




#14 N-DOTRIACONTANE

R.T.: 18.948 min
Delta R.T.: 0.002 min
Response: 28468
Conc: 0.22 ug/ml




#15 N-TETRATRIACONTANE

R.T.: 19.738 min
Delta R.T.: 0.008 min
Response: 80363
Conc: 0.72 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.474 min
Delta R.T.: 0.006 min
Response: 122933
Conc: 1.20 ug/ml



#18 N-TETRACONTANE

R.T.: 22.179 min
Delta R.T.: 0.047 min
Response: 87592
Conc: 0.85 ug/ml

Instrument :
FID_F
ClientSampleId :