

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF021723\
 Data File : FF012143.D
 Signal(s) : FID2B.ch
 Acq On : 17 Feb 2023 09:57
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 17 22:04:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF012723.M
 Quant Title :
 QLast Update : Fri Jan 27 14:21:25 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.064	6350505	55.111 ug/ml
Target Compounds			
1) N-OCTANE	2.081	5669505	49.282 ug/ml
2) N-DECANE	4.610	6671478	56.585 ug/ml
3) N-DODECANE	6.776	6814695	56.888 ug/ml
4) N-TETRADECANE	8.603	6992062	58.061 ug/ml
5) N-HEXADECANE	10.211	7083516	56.788 ug/ml
6) N-OCTADECANE	11.655	7228995	56.853 ug/ml
7) N-EICOSANE	12.965	7335354	56.837 ug/ml
8) N-DOCOSANE	14.164	7281847	57.125 ug/ml
10) N-TETRACOSANE	15.269	7165974	55.070 ug/ml
11) N-HEXACOSANE	16.292	6978483	53.621 ug/ml
12) N-OCTACOSANE	17.242	6937415	53.350 ug/ml
13) N-TRIACONTANE	18.129	7017757	54.377 ug/mlm
14) N-DOTRIACONTANE	18.962	6873267	52.812 ug/mlm
15) N-TETRATRIACONTANE	19.746	6309227	56.670 ug/mlm
16) N-HEXATRIACONTANE	20.486	5941923	57.844 ug/mlm
17) N-OCTATRIACONTANE	21.232	5371915	54.534 ug/mlm
18) N-TETRACONTANE	22.161	4824566	46.552 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF021723\
Data File : FF012143.D
Signal(s) : FID2B.ch
Acq On : 17 Feb 2023 09:57
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 17 22:04:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF012723.M
Quant Title :
QLast Update : Fri Jan 27 14:21:25 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

