

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF021723\
 Data File : FF012154.D
 Signal(s) : FID2B.ch
 Acq On : 17 Feb 2023 20:31
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 17 22:05:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF012723.M
 Quant Title :
 QLast Update : Fri Jan 27 14:21:25 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.059	5704460	49.504 ug/ml
Target Compounds			
1) N-OCTANE	2.081	5481236	47.646 ug/ml
2) N-DECANE	4.608	6403675	54.313 ug/ml
3) N-DODECANE	6.773	6497807	54.243 ug/ml
4) N-TETRADECANE	8.600	6604317	54.842 ug/ml
5) N-HEXADECANE	10.208	6593490	52.860 ug/ml
6) N-OCTADECANE	11.651	6621647	52.076 ug/ml
7) N-EICOSANE	12.961	6643235	51.474 ug/ml
8) N-DOCOSANE	14.160	6549493	51.380 ug/ml
10) N-TETRACOSANE	15.263	6451978	49.583 ug/ml
11) N-HEXACOSANE	16.285	6302327	48.426 ug/ml
12) N-OCTACOSANE	17.235	6308024	48.510 ug/ml
13) N-TRIACONTANE	18.124	6386891	49.489 ug/mlm
14) N-DOTRIACONTANE	18.956	6217311	47.772 ug/ml
15) N-TETRATRIACONTANE	19.740	5745080	51.602 ug/ml
16) N-HEXATRIACONTANE	20.480	5196686	50.589 ug/mlm
17) N-OCTATRIACONTANE	21.224	4575697	46.451 ug/mlm
18) N-TETRACONTANE	22.152	4413876	42.590 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF021723\
Data File : FF012154.D
Signal(s) : FID2B.ch
Acq On : 17 Feb 2023 20:31
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 17 22:05:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF012723.M
Quant Title :
QLast Update : Fri Jan 27 14:21:25 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

