

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF021819\
 Data File : FF003552.D
 Signal(s) : FID2B.ch
 Acq On : 18 Feb 2019 18:54
 Operator : AJ\SJ
 Sample : K1518-07MSD
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 19 00:49:35 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
 Quant Title :
 QLast Update : Fri Feb 01 22:57:49 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.609	1986922	16.458 ug/ml
Target Compounds			
1) N-OCTANE	2.242	1612273	12.077 ug/ml
2) N-DECANE	4.336	2524572	18.864 ug/ml
3) N-DODECANE	6.357	2663289	19.753 ug/ml
4) N-TETRADECANE	8.151	2662680	19.672 ug/ml
5) N-HEXADECANE	9.751	2720870	19.966 ug/ml
6) N-OCTADECANE	11.193	2693062	19.572 ug/ml
7) N-EICOSANE	12.505	2671707	19.428 ug/ml
8) N-DOCOSANE	13.707	2629992	19.223 ug/ml
10) N-TETRACOSANE	14.813	2858399	21.019 ug/ml
11) N-HEXACOSANE	15.838	2613232	19.267 ug/ml
12) N-OCTACOSANE	16.792	2696134	19.692 ug/ml
13) N-TRIACONTANE	17.683	2677757	18.965 ug/ml
14) N-DOTRIACONTANE	18.518	2705032	19.601 ug/ml
15) N-TETRATRIACONTANE	19.305	2772890	21.862 ug/ml
16) N-HEXATRIACONTANE	20.049	2958514	25.309 ug/ml
17) N-OCTATRIACONTANE	20.786	3068328	28.385 ug/ml
18) N-TETRACONTANE	21.681	3322992	30.692 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF021819\
Data File : FF003552.D
Signal(s) : FID2B.ch
Acq On : 18 Feb 2019 18:54
Operator : AJ\SJ
Sample : K1518-07MSD
Misc :
ALS Vial : 62 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 19 00:49:35 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
Quant Title :
QLast Update : Fri Feb 01 22:57:49 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#1 N-OCTANE

R.T.: 2.242 min
Delta R.T.: -0.006 min
Response: 1612273
Conc: 12.08 ug/ml




#2 N-DECANE

R.T.: 4.336 min
Delta R.T.: -0.006 min
Response: 2524572
Conc: 18.86 ug/ml




#3 N-DODECANE

R.T.: 6.357 min
Delta R.T.: -0.004 min
Response: 2663289
Conc: 19.75 ug/ml




#4 N-TETRADECANE

R.T.: 8.151 min
Delta R.T.: -0.003 min
Response: 2662680
Conc: 19.67 ug/ml




#5 N-HEXADECANE

R.T.: 9.751 min
Delta R.T.: -0.002 min
Response: 2720870
Conc: 19.97 ug/ml




#6 N-OCTADECANE

R.T.: 11.193 min
Delta R.T.: -0.002 min
Response: 2693062
Conc: 19.57 ug/ml




#7 N-EICOSANE

R.T.: 12.505 min
Delta R.T.: -0.003 min
Response: 2671707
Conc: 19.43 ug/ml




#8 N-DOCOSANE

R.T.: 13.707 min
Delta R.T.: -0.001 min
Response: 2629992
Conc: 19.22 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.609 min
Delta R.T.: 0.000 min
Response: 1986922
Conc: 16.46 ug/ml




#10 N-TETRACOSANE

R.T.: 14.813 min
Delta R.T.: -0.001 min
Response: 2858399
Conc: 21.02 ug/ml




#11 N-HEXACOSANE

R.T.: 15.838 min
Delta R.T.: 0.000 min
Response: 2613232
Conc: 19.27 ug/ml




#12 N-OCTACOSANE

R.T.: 16.792 min
Delta R.T.: 0.000 min
Response: 2696134
Conc: 19.69 ug/ml




#13 N-TRIACONTANE

R.T.: 17.683 min
Delta R.T.: 0.000 min
Response: 2677757
Conc: 18.97 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.518 min
Delta R.T.: -0.002 min
Response: 2705032
Conc: 19.60 ug/ml




#15 N-TETRATRIACONTANE

R.T.: 19.305 min
Delta R.T.: -0.002 min
Response: 2772890
Conc: 21.86 ug/ml




#16 N-HEXATRIACONTANE

R.T.: 20.049 min
Delta R.T.: -0.001 min
Response: 2958514
Conc: 25.31 ug/ml



#17 N-OCTATRIACONTANE

R.T.: 20.786 min
Delta R.T.: -0.001 min
Response: 3068328
Conc: 28.39 ug/ml



#18 N-TETRACONTANE

R.T.: 21.681 min
Delta R.T.: -0.003 min
Response: 3322992
Conc: 30.69 ug/ml