

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF021819\
 Data File : FF003555.D
 Signal(s) : FID2B.ch
 Acq On : 18 Feb 2019 20:23
 Operator : AJ\SJ
 Sample : K1518-13
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 19 00:50:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
 Quant Title :
 QLast Update : Fri Feb 01 22:57:49 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.611	2497870	20.690 ug/ml
Target Compounds			
3) N-DODECANE	6.360	437161	3.242 ug/ml
4) N-TETRADECANE	8.147	100109	0.740 ug/ml
7) N-EICOSANE	12.507	82264	0.598 ug/ml
8) N-DOCOSANE	13.728	380116	2.778 ug/ml
10) N-TETRACOSANE	14.787	2245863	16.514 ug/ml
11) N-HEXACOSANE	15.839	809725	5.970 ug/ml
12) N-OCTACOSANE	16.798	2204920	16.104 ug/ml
13) N-TRIACONTANE	17.660	5261111	37.262 ug/ml
14) N-DOTRIACONTANE	18.512	815297	5.908 ug/ml
15) N-TETRATRIACONTANE	19.356	79178	0.624 ug/ml
16) N-HEXATRIACONTANE	20.091	171268	1.465 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF021819\
Data File : FF003555.D
Signal(s) : FID2B.ch
Acq On : 18 Feb 2019 20:23
Operator : AJ\SJ
Sample : K1518-13
Misc :
ALS Vial : 65 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 19 00:50:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
Quant Title :
QLast Update : Fri Feb 01 22:57:49 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE

R.T.: 6.360 min
Delta R.T.: 0.000 min
Response: 437161
Conc: 3.24 ug/ml




#4 N-TETRADECANE

R.T.: 8.147 min
Delta R.T.: -0.006 min
Response: 100109
Conc: 0.74 ug/ml




#7 N-EICOSANE

R.T.: 12.507 min
Delta R.T.: -0.001 min
Response: 82264
Conc: 0.60 ug/ml




#8 N-DOCOSANE

R.T.: 13.728 min
Delta R.T.: 0.020 min
Response: 380116
Conc: 2.78 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.611 min
Delta R.T.: 0.001 min
Response: 2497870
Conc: 20.69 ug/ml




#10 N-TETRACOSANE

R.T.: 14.787 min
Delta R.T.: -0.027 min
Response: 2245863
Conc: 16.51 ug/ml




#11 N-HEXACOSANE

R.T.: 15.839 min
Delta R.T.: 0.000 min
Response: 809725
Conc: 5.97 ug/ml




#12 N-OCTACOSANE

R.T.: 16.798 min
Delta R.T.: 0.006 min
Response: 2204920
Conc: 16.10 ug/ml




#13 N-TRIACONTANE

R.T.: 17.660 min
Delta R.T.: -0.023 min
Response: 5261111
Conc: 37.26 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.512 min
Delta R.T.: -0.007 min
Response: 815297
Conc: 5.91 ug/ml



#15 N-TETRATRIACONTANE

R.T.: 19.356 min
Delta R.T.: 0.049 min
Response: 79178
Conc: 0.62 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.091 min
Delta R.T.: 0.041 min
Response: 171268
Conc: 1.47 ug/ml