

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF022023\
 Data File : FF012157.D
 Signal(s) : FID2B.ch
 Acq On : 20 Feb 2023 09:31
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 21 01:21:35 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF012723.M
 Quant Title :
 QLast Update : Fri Jan 27 14:21:25 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.071	6371680	55.294 ug/ml
Target Compounds			
1) N-OCTANE	2.083	5676027	49.339 ug/ml
2) N-DECANE	4.613	6901476	58.536 ug/ml
3) N-DODECANE	6.779	7039799	58.768 ug/ml
4) N-TETRADECANE	8.607	7199664	59.785 ug/ml
5) N-HEXADECANE	10.216	7259770	58.201 ug/ml
6) N-OCTADECANE	11.660	7361605	57.896 ug/ml
7) N-EICOSANE	12.972	7431919	57.585 ug/ml
8) N-DOCOSANE	14.171	7322311	57.442 ug/ml
10) N-TETRACOSANE	15.276	7169882	55.100 ug/ml
11) N-HEXACOSANE	16.298	6974834	53.593 ug/ml
12) N-OCTACOSANE	17.248	6965023	53.562 ug/ml
13) N-TRIACONTANE	18.137	7106258	55.063 ug/mlm
14) N-DOTRIACONTANE	18.971	7299392	56.087 ug/ml
15) N-TETRATRIACONTANE	19.755	6567368	58.988 ug/mlm
16) N-HEXATRIACONTANE	20.495	6061317	59.007 ug/mlm
17) N-OCTATRIACONTANE	21.243	5549462	56.336 ug/mlm
18) N-TETRACONTANE	22.180	4993365	48.181 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF022023\
Data File : FF012157.D
Signal(s) : FID2B.ch
Acq On : 20 Feb 2023 09:31
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 21 01:21:35 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF012723.M
Quant Title :
QLast Update : Fri Jan 27 14:21:25 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

