

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF022223\
 Data File : FF012190.D
 Signal(s) : FID2B.ch
 Acq On : 22 Feb 2023 17:22
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/23/2023
 Supervised By :Ankita Jodhani 02/23/2023

Integration File: autoint1.e
 Quant Time: Feb 23 00:50:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF022123.M
 Quant Title :
 QLast Update : Tue Feb 21 16:12:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.041	5098638	42.975 ug/ml
Target Compounds			
1) N-OCTANE	2.079	5647993	42.408 ug/ml
2) N-DECANE	4.599	5738901	42.798 ug/ml
3) N-DODECANE	6.762	5860572	43.137 ug/ml
4) N-TETRADECANE	8.588	5928625	42.587 ug/ml
5) N-HEXADECANE	10.194	6048523	42.491 ug/ml
6) N-OCTADECANE	11.636	6065392	42.667 ug/ml
7) N-EICOSANE	12.946	6012873	42.619 ug/ml
8) N-DOCOSANE	14.143	5767752	42.758 ug/ml
10) N-TETRACOSANE	15.245	5697930	43.008 ug/ml
11) N-HEXACOSANE	16.265	5533529	43.379 ug/ml
12) N-OCTACOSANE	17.215	5368637	43.740 ug/ml
13) N-TRIACONTANE	18.102	5138809	44.020 ug/ml
14) N-DOTRIACONTANE	18.933	4841110	43.479 ug/mlm
15) N-TETRATRIACONTANE	19.716	4279456	43.251 ug/mlm
16) N-HEXATRIACONTANE	20.456	3866436	40.770 ug/ml
17) N-OCTATRIACONTANE	21.194	3523817	41.098 ug/mlm
18) N-TETRACONTANE	22.110	3577579	39.289 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF022223\
Data File : FF012190.D
Signal(s) : FID2B.ch
Acq On : 22 Feb 2023 17:22
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/23/2023
Supervised By :Ankita Jodhani 02/23/2023

Integration File: autoint1.e
Quant Time: Feb 23 00:50:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF022123.M
Quant Title :
QLast Update : Tue Feb 21 16:12:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

