

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF022323\
Data File : FF012217.D
Signal(s) : FID2B.ch
Acq On : 23 Feb 2023 14:57
Operator : YP\AJ
Sample : PB151004BL
Misc :
ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 23 16:04:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF022123.M
Quant Title :
QLast Update : Tue Feb 21 16:12:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.034	2373768	20.008 ug/ml
Target Compounds			
13) N-TRIACONTANE	18.095	56775	0.486 ug/ml
14) N-DOTRIACONTANE	18.928	89355	0.803 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF022323\
Data File : FF012217.D
Signal(s) : FID2B.ch
Acq On : 23 Feb 2023 14:57
Operator : YP\AJ
Sample : PB151004BL
Misc :
ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 23 16:04:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF022123.M
Quant Title :
QLast Update : Tue Feb 21 16:12:39 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 15.034 min
Delta R.T.: -0.026 min
Response: 2373768
Conc: 20.01 ug/ml



#13 N-TRIACONTANE

R.T.: 18.095 min
Delta R.T.: -0.028 min
Response: 56775
Conc: 0.49 ug/ml



#14 N-DOTRIACONTANE

R.T.: 18.928 min
Delta R.T.: -0.026 min
Response: 89355
Conc: 0.80 ug/ml