

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF022420\
 Data File : FF006466.D
 Signal(s) : FID2B.ch
 Acq On : 24 Feb 2020 16:19
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :

Manual Integrations
 APPROVED

Sohil
 2/25/2020 10:32:13 AM

Integration File: autoint1.e
 Quant Time: Feb 25 00:30:07 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF021020.M
 Quant Title :
 QLast Update : Tue Feb 11 10:27:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.606	6680002	54.831 ug/ml
Target Compounds			
1) N-OCTANE	2.282	7642781	57.634 ug/ml
2) N-DECANE	4.372	8209273	61.471 ug/ml
3) N-DODECANE	6.387	8249978	60.363 ug/ml
4) N-TETRADECANE	8.178	8131699	59.118 ug/ml
5) N-HEXADECANE	9.774	8153625	58.115 ug/ml
6) N-OCTADECANE	11.211	8072465	57.392 ug/ml
7) N-EICOSANE	12.517	7937967	56.324 ug/ml
8) N-DOCOSANE	13.711	7782118	55.373 ug/ml
10) N-TETRACOSANE	14.810	7699813	55.104 ug/ml
11) N-HEXACOSANE	15.826	7394871	53.566 ug/ml
12) N-OCTACOSANE	16.773	7350731	53.263 ug/ml
13) N-TRIACONTANE	17.657	7372924	53.315 ug/ml
14) N-DOTRIACONTANE	18.485	7052283	53.604 ug/ml
15) N-TETRATRIACONTANE	19.265	6164898	54.895 ug/ml
16) N-HEXATRIACONTANE	20.001	5206246	53.795 ug/mlm
17) N-OCTATRIACONTANE	20.722	4072917	57.679 ug/ml
18) N-TETRACONTANE	21.596	4042997	57.275 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF022420\
Data File : FF006466.D
Signal(s) : FID2B.ch
Acq On : 24 Feb 2020 16:19
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampled :

Manual Integrations
APPROVED

Sohil
2/25/2020 10:32:13 AM

Integration File: autoint1.e
Quant Time: Feb 25 00:30:07 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF021020.M
Quant Title :
QLast Update : Tue Feb 11 10:27:03 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

