

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF022619\
 Data File : FF003645.D
 Signal(s) : FID2B.ch
 Acq On : 26 Feb 2019 15:25
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Feb 27 00:03:21 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
 Quant Title :
 QLast Update : Fri Feb 01 22:57:49 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.611	5827230	48.268 ug/ml
Target Compounds			
1) N-OCTANE	2.235	7514635	56.287 ug/ml
2) N-DECANE	4.333	7722156	57.701 ug/ml
3) N-DODECANE	6.356	7677681	56.944 ug/ml
4) N-TETRADECANE	8.149	7534815	55.667 ug/ml
5) N-HEXADECANE	9.751	7334142	53.818 ug/ml
6) N-OCTADECANE	11.194	7115578	51.712 ug/ml
7) N-EICOSANE	12.506	6852293	49.829 ug/ml
8) N-DOCOSANE	13.708	6647665	48.588 ug/ml
10) N-TETRACOSANE	14.815	6551233	48.173 ug/ml
11) N-HEXACOSANE	15.840	6490101	47.850 ug/ml
12) N-OCTACOSANE	16.794	6534585	47.727 ug/ml
13) N-TRIACONTANE	17.684	6577423	46.585 ug/ml
14) N-DOTRIACONTANE	18.519	6413441	46.473 ug/ml
15) N-TETRATRIACONTANE	19.309	6269181	49.427 ug/mlm
16) N-HEXATRIACONTANE	20.053	6091961	52.114 ug/ml
17) N-OCTATRIACONTANE	20.790	5584420	51.661 ug/ml
18) N-TETRACONTANE	21.682	5512513	50.915 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF022619\
Data File : FF003645.D
Signal(s) : FID2B.ch
Acq On : 26 Feb 2019 15:25
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Feb 27 00:03:21 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF020119.M
Quant Title :
QLast Update : Fri Feb 01 22:57:49 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

