

Method Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\
 Method File : FF030119.M
 Title :
 Last Update : Mon Mar 04 08:35:38 2019
 Response Via : Initial Calibration

Calibration Files

50 =FF003671.D 100 =FF003670.D 20 =FF003672.D
 10 =FF003673.D 5 =FF003674.D

Compound		50	100	20	10	5	Avg	%RSD	
1)	N-OCTANE	1.386	1.464	1.499	1.544	1.783	1.535	E5	9.79
2)	N-DECANE	1.401	1.478	1.508	1.545	1.759	1.538	E5	8.74
3)	N-DODECANE	1.428	1.505	1.528	1.562	1.780	1.561	E5	8.47
4)	N-TETRADECANE	1.468	1.523	1.570	1.614	1.842	1.603	E5	8.99
5)	N-HEXADECANE	1.472	1.533	1.598	1.649	1.892	1.629	E5	9.93
6)	N-OCTADECANE	1.483	1.534	1.611	1.674	1.934	1.648	E5	10.69
7)	N-EICOSANE	1.466	1.513	1.600	1.673	1.928	1.636	E5	11.10
8)	N-DOCOSANE	1.446	1.486	1.577	1.654	1.906	1.614	E5	11.30
9) S	TETRACOSANE-d50 (S...	1.250	1.279	1.361	1.429	1.642	1.392	E5	11.24
10)	N-TETRACOSANE	1.419	1.453	1.544	1.625	1.870	1.582	E5	11.39
11)	N-HEXACOSANE	1.387	1.416	1.507	1.589	1.819	1.543	E5	11.22
12)	N-OCTACOSANE	1.351	1.373	1.471	1.557	1.805	1.511	E5	12.14
13)	N-TRIACONTANE	1.312	1.329	1.426	1.511	1.776	1.471	E5	12.82
14)	N-DOTRIACONTANE	1.258	1.266	1.358	1.486	1.743	1.422	E5	14.18
15)	N-TETRATRIACONTANE	1.203	1.221	1.294	1.361	1.616	1.339	E5	12.47
16)	N-HEXATRIACONTANE	1.174	1.196	1.243	1.300	1.612	1.305	E5	13.67
17)	N-OCTATRIACONTANE	1.045	1.076	1.111	1.160	1.381	1.155	E5	11.59
18)	N-TETRACONTANE	1.117	1.132	1.104	1.425	1.443	1.244	E5	13.97

(#) = Out of Range