

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030421\
 Data File : FF008157.D
 Signal(s) : FID2B.ch
 Acq On : 04 Mar 2021 19:39
 Operator : DD\AJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 10 TRPH STD

Manual Integrations
 APPROVED

Ankita
 3/5/2021 9:34:01 AM

Integration File: autoint1.e
 Quant Time: Mar 05 02:06:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030421.M
 Quant Title :
 QLast Update : Fri Mar 05 02:04:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.668	1497581	10.998 ug/ml
Target Compounds			
1) N-OCTANE	2.296	1319419	10.308 ug/ml
2) N-DECANE	4.396	1346491	10.121 ug/ml
3) N-DODECANE	6.418	1420727	10.274 ug/ml
4) N-TETRADECANE	8.213	1541335	10.565 ug/ml
5) N-HEXADECANE	9.814	1591931	10.745 ug/ml
6) N-OCTADECANE	11.257	1655759	10.886 ug/ml
7) N-EICOSANE	12.568	1676376	10.969 ug/ml
8) N-DOCOSANE	13.767	1687273	10.993 ug/ml
10) N-TETRACOSANE	14.872	1699303	11.028 ug/ml
11) N-HEXACOSANE	15.894	1680079	11.039 ug/ml
12) N-OCTACOSANE	16.845	1667072	11.055 ug/ml
13) N-TRIACONTANE	17.735	1629251	11.098 ug/ml
14) N-DOTRIACONTANE	18.568	1567266	11.173 ug/ml
15) N-TETRATRIACONTANE	19.352	1468257	11.058 ug/ml
16) N-HEXATRIACONTANE	20.093	1366167	11.040 ug/mlm
17) N-OCTATRIACONTANE	20.835	1253618	10.985 ug/ml
18) N-TETRACONTANE	21.744	1298454	11.165 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030421\
Data File : FF008157.D
Signal(s) : FID2B.ch
Acq On : 04 Mar 2021 19:39
Operator : DD\AJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 61 Sample Multiplier: 1

Instrument :
FID_F
ClientSampled :
10 TRPH STD

Manual Integrations
APPROVED

Ankita
3/5/2021 9:34:01 AM

Integration File: autoint1.e
Quant Time: Mar 05 02:06:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030421.M
Quant Title :
QLast Update : Fri Mar 05 02:04:51 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

