

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030520\
 Data File : FF006495.D
 Signal(s) : FID2B.ch
 Acq On : 05 Mar 2020 12:10
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 06 00:31:46 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF021020.M
 Quant Title :
 QLast Update : Tue Feb 11 10:27:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.592	6255965	51.350 ug/ml
Target Compounds			
2) N-DECANE	4.354	8295141	62.114 ug/ml
3) N-DODECANE	6.370	8266478	60.484 ug/ml
4) N-TETRADECANE	8.162	8039163	58.445 ug/ml
5) N-HEXADECANE	9.759	7941124	56.601 ug/ml
6) N-OCTADECANE	11.197	7769775	55.240 ug/ml
7) N-EICOSANE	12.502	7540518	53.504 ug/ml
8) N-DOCOSANE	13.695	7345777	52.268 ug/ml
10) N-TETRACOSANE	14.795	7402728	52.978 ug/ml
11) N-HEXACOSANE	15.812	6944161	50.302 ug/ml
12) N-OCTACOSANE	16.758	6876017	49.824 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030520\
Data File : FF006495.D
Signal(s) : FID2B.ch
Acq On : 05 Mar 2020 12:10
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 06 00:31:46 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF021020.M
Quant Title :
QLast Update : Tue Feb 11 10:27:03 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

