

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030521\
 Data File : FF008162.D
 Signal(s) : FID2B.ch
 Acq On : 05 Mar 2021 09:21
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Mar 05 22:01:21 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030421.M
 Quant Title :
 QLast Update : Fri Mar 05 09:17:37 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.671	6818908	48.286 ug/ml
Target Compounds			
2) N-DECANE	4.397	6725134	51.491 ug/ml
3) N-DODECANE	6.419	6964697	50.928 ug/ml
4) N-TETRADECANE	8.215	7307530	49.680 ug/ml
5) N-HEXADECANE	9.816	7390764	48.928 ug/ml
6) N-OCTADECANE	11.259	7587343	48.443 ug/ml
7) N-EICOSANE	12.570	7635744	48.212 ug/ml
8) N-DOCOSANE	13.770	7682618	48.236 ug/ml
10) N-TETRACOSANE	14.875	7739543	48.294 ug/ml
11) N-HEXACOSANE	15.898	7666174	48.454 ug/ml
12) N-OCTACOSANE	16.849	7635251	48.618 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030521\
Data File : FF008162.D
Signal(s) : FID2B.ch
Acq On : 05 Mar 2021 09:21
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Mar 05 22:01:21 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030421.M
Quant Title :
QLast Update : Fri Mar 05 09:17:37 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

