

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030521\
 Data File : FF008184.D
 Signal(s) : FID2B.ch
 Acq On : 05 Mar 2021 21:56
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Mar 05 23:24:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030421.M
 Quant Title :
 QLast Update : Fri Mar 05 09:17:37 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.670	6788128	48.068 ug/ml
Target Compounds			
2) N-DECANE	4.396	6604506	50.567 ug/ml
3) N-DODECANE	6.419	6879139	50.302 ug/ml
4) N-TETRADECANE	8.215	7262109	49.371 ug/ml
5) N-HEXADECANE	9.816	7386985	48.903 ug/ml
6) N-OCTADECANE	11.259	7599919	48.524 ug/ml
7) N-EICOSANE	12.570	7648021	48.289 ug/ml
8) N-DOCOSANE	13.769	7668804	48.149 ug/ml
10) N-TETRACOSANE	14.875	7672888	47.878 ug/ml
11) N-HEXACOSANE	15.897	7532962	47.612 ug/ml
12) N-OCTACOSANE	16.848	7428438	47.301 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030521\
Data File : FF008184.D
Signal(s) : FID2B.ch
Acq On : 05 Mar 2021 21:56
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Mar 05 23:24:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030421.M
Quant Title :
QLast Update : Fri Mar 05 09:17:37 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

