

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030619\
 Data File : FF003742.D
 Signal(s) : FID2B.ch
 Acq On : 06 Mar 2019 11:44
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Mar 07 00:19:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030119.M
 Quant Title :
 QLast Update : Mon Mar 04 08:35:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.596	6207278	44.584 ug/ml
Target Compounds			
2) N-DECANE	4.313	8230907	53.510 ug/ml
3) N-DODECANE	6.337	8226148	52.710 ug/ml
4) N-TETRADECANE	8.132	8123952	50.669 ug/ml
5) N-HEXADECANE	9.735	8050761	49.430 ug/ml
6) N-OCTADECANE	11.180	7853363	47.667 ug/ml
7) N-EICOSANE	12.493	7546991	46.133 ug/ml
8) N-DOCOSANE	13.694	7255020	44.953 ug/ml
10) N-TETRACOSANE	14.801	7019061	44.362 ug/ml
11) N-HEXACOSANE	15.826	6848161	44.370 ug/ml
12) N-OCTACOSANE	16.779	6797647	44.984 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030619\
Data File : FF03742.D
Signal(s) : FID2B.ch
Acq On : 06 Mar 2019 11:44
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Mar 07 00:19:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030119.M
Quant Title :
QLast Update : Mon Mar 04 08:35:38 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

