

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030619\
 Data File : FF003746.D
 Signal(s) : FID2B.ch
 Acq On : 06 Mar 2019 14:44
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Mar 07 00:20:42 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030119.M
 Quant Title :
 QLast Update : Mon Mar 04 08:35:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.596	6342580	45.555 ug/ml
Target Compounds			
2) N-DECANE	4.322	8617862	56.026 ug/ml
3) N-DODECANE	6.344	8579430	54.974 ug/ml
4) N-TETRADECANE	8.137	8509363	53.072 ug/ml
5) N-HEXADECANE	9.739	8330739	51.149 ug/ml
6) N-OCTADECANE	11.182	8104065	49.188 ug/ml
7) N-EICOSANE	12.494	7752521	47.389 ug/ml
8) N-DOCOSANE	13.695	7416790	45.956 ug/ml
10) N-TETRACOSANE	14.801	7142538	45.143 ug/ml
11) N-HEXACOSANE	15.824	6956833	45.074 ug/ml
12) N-OCTACOSANE	16.779	6884317	45.557 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030619\
Data File : FF03746.D
Signal(s) : FID2B.ch
Acq On : 06 Mar 2019 14:44
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Mar 07 00:20:42 2019
Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030119.M
Quant Title :
QLast Update : Mon Mar 04 08:35:38 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

