

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030620\
 Data File : FF006503.D
 Signal(s) : FID2B.ch
 Acq On : 06 Mar 2020 10:12
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 06 22:55:47 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF021020.M
 Quant Title :
 QLast Update : Tue Feb 11 10:27:03 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.593	5794325	47.561 ug/ml
Target Compounds				
2)	N-DECANE	4.364	7452158	55.802 ug/ml
3)	N-DODECANE	6.378	7541207	55.177 ug/ml
4)	N-TETRADECANE	8.168	7437556	54.072 ug/ml
5)	N-HEXADECANE	9.764	7402270	52.760 ug/ml
6)	N-OCTADECANE	11.200	7241647	51.486 ug/ml
7)	N-EICOSANE	12.505	7038668	49.943 ug/ml
8)	N-DOCOSANE	13.698	6805919	48.427 ug/ml
10)	N-TETRACOSANE	14.797	6679136	47.799 ug/ml
11)	N-HEXACOSANE	15.813	6327797	45.837 ug/ml
12)	N-OCTACOSANE	16.759	6201862	44.939 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030620\
Data File : FF006503.D
Signal(s) : FID2B.ch
Acq On : 06 Mar 2020 10:12
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 06 22:55:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF021020.M
Quant Title :
QLast Update : Tue Feb 11 10:27:03 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

