

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030821\
 Data File : FF008187.D
 Signal(s) : FID2B.ch
 Acq On : 08 Mar 2021 09:39
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 09 00:36:19 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030421.M
 Quant Title :
 QLast Update : Fri Mar 05 09:17:37 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.668	6428178	45.519 ug/ml
Target Compounds			
2) N-DECANE	4.397	6506089	49.814 ug/ml
3) N-DODECANE	6.419	6810150	49.798 ug/ml
4) N-TETRADECANE	8.214	7105131	48.304 ug/ml
5) N-HEXADECANE	9.815	7231228	47.872 ug/ml
6) N-OCTADECANE	11.257	7423222	47.395 ug/ml
7) N-EICOSANE	12.569	7608423	48.039 ug/ml
8) N-DOCOSANE	13.767	7599414	47.714 ug/ml
10) N-TETRACOSANE	14.872	7487760	46.723 ug/ml
11) N-HEXACOSANE	15.894	7341068	46.399 ug/ml
12) N-OCTACOSANE	16.845	7191784	45.794 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF030821\
Data File : FF008187.D
Signal(s) : FID2B.ch
Acq On : 08 Mar 2021 09:39
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 09 00:36:19 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF030421.M
Quant Title :
QLast Update : Fri Mar 05 09:17:37 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

