

rteres

Instrument :

FID_F

LabSampleId :

5 TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/11/2022

Supervised By :Ankita Jodhani 03/11/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031022
Data File : FF010020.D
Signal(s) : FID2B.ch
Acq On : 10 Mar 2022 12:16
Sample : 5 TRPH STD
Misc :
ALS Vial : 57 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031022.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.335	2.305	2.385	BB	134849	993790	91.37%	5.922%
2	4.434	4.402	4.468	BB	112547	1026337	94.36%	6.116%
3	6.452	6.415	6.492	BB	105661	1037333	95.37%	6.182%
4	8.245	8.212	8.308	BB	102356	1049209	96.46%	6.252%
5	9.846	9.798	9.892	BB	98619	1055290	97.02%	6.289%
6	11.287	11.225	11.322	BB	97255	1087687	100.00%	6.482%
7	12.596	12.492	12.652	BB	89890	1061335	97.58%	6.325%
8	13.794	13.762	13.838	VB	85944	1041446	95.75%	6.206%
9	14.694	14.655	14.724	BV	65297	871673	80.14%	5.194%
10	14.898	14.862	14.958	BB	83979	1023941	94.14%	6.102%
11	15.920	15.868	15.962	BB	79081	1006833	92.57%	6.000%
12	16.869	16.805	16.928	BB	73968	971161	89.29%	5.787%
13	17.757	17.708	17.802	BB	65255	930528	85.55%	5.545%
14	18.590	18.522	18.645	BB	62694	883826	81.26%	5.267%
15	19.374	19.292	19.412	BB	56809	808249	74.31%	4.816%
16	20.114	20.062	20.149	BV	49537	723346	66.50%	4.310%
17	20.853	20.768	20.908	BB	35998	643081	59.12%	3.832%
18	21.765	21.675	21.855	BB	23919	566120	52.05%	3.374%
Sum of corrected areas:						16781184		

FF031022.M Fri Mar 11 09:06:04 2022