

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031120\
 Data File : FF006534.D
 Signal(s) : FID2B.ch
 Acq On : 11 Mar 2020 14:30
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 12 01:10:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031020.M
 Quant Title :
 QLast Update : Tue Mar 10 14:59:43 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.588	6487066	50.404 ug/ml
Target Compounds			
2) N-DECANE	4.360	7180999	52.169 ug/ml
3) N-DODECANE	6.374	7363157	52.206 ug/ml
4) N-TETRADECANE	8.163	7395167	51.708 ug/ml
5) N-HEXADECANE	9.759	7520548	51.305 ug/ml
6) N-OCTADECANE	11.195	7558070	51.149 ug/ml
7) N-EICOSANE	12.500	7540234	50.733 ug/ml
8) N-DOCOSANE	13.693	7488702	50.423 ug/ml
10) N-TETRACOSANE	14.791	7406934	50.116 ug/ml
11) N-HEXACOSANE	15.807	7224300	49.851 ug/ml
12) N-OCTACOSANE	16.751	7108296	49.630 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF031120\
Data File : FF006534.D
Signal(s) : FID2B.ch
Acq On : 11 Mar 2020 14:30
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 12 01:10:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF031020.M
Quant Title :
QLast Update : Tue Mar 10 14:59:43 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

